



Ohio Administrative Code

Rule 3701:1-58-24 Calibration of survey instruments.

Effective: October 4, 2010

(A) A licensee shall calibrate the survey instruments used to show compliance with this chapter and Chapter 3701:1-38 of the Administrative Code before first use, annually, and following a repair that affects the calibration. A licensee shall:

(1) Calibrate all scales with readings up to ten millisievert (one thousand millirem) per hour with a radiation source;

(2) Calibrate two separate readings on each scale or decade that will be used to show compliance;
and

(3) Conspicuously note on the instrument the date of calibration.

(B) A licensee may not use survey instruments if the difference between the indicated exposure rate and the calculated exposure rate is more than twenty per cent.

(C) A licensee shall retain a record of each survey instrument calibration in accordance with rule 3701:1-58-78 of the Administrative Code.
